

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Rogowski current sensor

Name and address of the applicant

ALGODUE ELETTRONICA SRL
VIA P. GOBETTI 16/F
28014 MAGGIORA NO
Italy

Name and address of the manufacturer

ALGODUE ELETTRONICA SRL
VIA P. GOBETTI 16/F
28014 MAGGIORA NO
Italy

Name and address of the factory

Note: When more than one factory, please report on page 2

☐ Additional information on page 2
ALGODUE ELETTRONICA SRL
VIA P. GOBETTI 16/F
28014 MAGGIORA NO
Italy

Ratings and principal characteristics

3.3 Vd.c. - 10 mA

Trademark (if any)

SOCOMECH

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

TF-XXX Series
Models TF-40 - TF-80 - TF-120 - TF-200 - TF-300 - TF-600 (see Test Report)

Additional information (if necessary may also be reported on page 2)

☐ Additional information on page 2
This CBTC is a modification to CBTC no. IT-18930 dated 2018-04-23 and they must be read in conjunction. Modification introduced: adding of new model TF-40 and modification of applicant and manufacturer address.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016, IEC 61010-2-032:2012

National differences:

CA, US

As shown in the Test Report Ref. No. which forms part of this Certificate

AE20-0059038-02

This CB Test Certificate is issued by the National Certification Body

IMQ S.p.A.
Via Quintiliano 43, IT-20138 Milano, Italy